

DECLARATION OF CONFORMITY

We, Bortec Biomedical Ltd.
#225, 604 – 1st Street SW
Calgary, Alberta
Canada, T2P 1M7

declare under sole responsibility that the product:

AMT-8 Multichannel EMG wire telemetry system

to which this declaration relates, are in conformity with the following standards and /or other normative documents:

Class: I
non sterile, non measuring

Safety: IEC 60601-1 with Amendments 1 & 2
C22.2 No. 601.1-M90
UL Std No. 2601-1

EMC: EN 55011 Class B
EN 60601-1-2
EN 60601-2-40


We hereby declare that all essential EMI/EMC and safety tests have been carried out and that the above named product is in conformity to all essential requirements of Medical Device Directive 93/42/EEC.

The Technical File documentation relevant to the above equipment is held at:

Dr. Uwe Windhorst
Ludvig-Prandtl Str 4
Weende-Nord, Goettingen
37077
Germany

Dr. Scott Day
President and CEO

Mr. Boris Kacmar
VP Product Development



Calgary
March 31, 2003
Calgary
March 31, 2003